




OCT Validation Phantom

APL-OP01



- Four patterns give comprehensive assessment of your OCT instrumentation.
- Patterns laser etched for long term resistance against drift.
- Small, stable and easy to transport for simple machine to machine repeatability testing.
- Gives accurate measurements of OCT performance in all 3 dimensions.




The APL-OP01 OCT phantom is a device to qualitatively and quantitatively characterize the performance of an OCT system, including spatially varying resolution, sensitivity, distortion and scan linearity. The phantom contains four geometrical patterns inscribed within a substrate to form localized refractive index changes.

The 4 test patterns are from left to right:

- Point Spread Function
- Sensitivity
- Lateral resolution
- Distortion

Each of the 4 patterns has different features but they all contain 8 layers. In each pattern, the first layer is located $100\mu\text{m}$ below the surface of the substrate and the separation between each subsequent layer is $75\mu\text{m}$. So the bottom layer is $625\mu\text{m}$ from the top surface of the substrate.

OCT Test Patterns



Note: Indexing lines are indicated in red.

OCT Validation Phantom

APL-OP01

Distortion Patterns

The Distortion pattern consists of $100\ \mu\text{m} \times 100\ \mu\text{m}$ grids.

Lateral Resolution Pattern

The purpose of the Lateral Resolution pattern is to measure the line spacing. Each line (n) is separated from the next line laterally ($n+1$) by a distance starting at $1\ \mu\text{m}$ and increasing in increments of $1\ \mu\text{m}$, up to a maximum of $110\ \mu\text{m}$.

The lines are separated into groups; each group contains 11 lines (11 spaces), with index lines (shown in red) separating each group. The following formula can be used to calculate line spacing:

$$\text{Line spacing} = 11(m-1)+n$$

m = the group number,
 n = the number of the line within the group




Diagram of the lateral resolution pattern with the group numbers above the line number within a group below

e.g. to calculate the line spacing for lines 7 and 8 in group 5 – the first value is 7 ($= n$) and the second value is 44 ($= 11m-1$) so the line spacing is $51\ \mu\text{m}$.

Lateral Resolution Pattern

The Sensitivity pattern is made up of pairs of lines (2 lines). Both lines in a pair have the same intensity but the intensity of each line pair is greater than that of the previous line pair.

The line pairs are grouped into sets. Each set consists of five line pairs. Each set, n , is separated by n and $n+1$ index lines *e.g.* set 3 has 3 index lines on one side and 4 index lines on the other.

Each layer consists of six sets of line pairs with five line pairs in each set (*i.e.* 30 pairs in total in each layer). The spacing between the lines is $150\ \mu\text{m}$ and each layer is offset laterally by $25\ \mu\text{m}$.

Point Spread Function Pattern

The Point Spread Function (PSF) pattern consists of a series of lines within each layer separated by $150\ \mu\text{m}$. Layers are laterally offset from the layer above by $25\ \mu\text{m}$.



OCT Validation Phantom

APL-OP01

Technical Specification

Environmental

Operational Temperature	0°C to +50°C
Humidity	5% – 95%, non-condensing

Physical

Mass	10 grams (in storage box 300 grams)
Dimensions	40 x 40 x 3 mm
Dimensions of Substrate	25 x 30 x 2 mm
Dimensions of Patterns	13 x 4 mm
Dimensional Tolerance	+/- 2 μm within zones A+B (diagram 1)
Substrate Material	Silica
Refractive index of substrate	1.45
Lateral Width of Lines	< 2 μm




Diagram left: For the Distortion Pattern the space within Zone A (dark blue hatched area) should be used. Within this area the inter line dimensional tolerance is +/- 0.8 μm . For the other patterns the space within Zone B (light orange area) and Zone A should be used. Within this area the inter line dimensional tolerance is +/- 0.8 μm . These tolerances are the 2 σ standard deviation of the inter-line spacing measurements.

Ordering Information

Product

APL-OP01	OCT Phantom APL-OP01 consists of 4 test patterns Point Spread Function, Sensitivity, Lateral resolution and Distortion. Supplied in wooden storage box.
----------	---

For North American sales enquiries call **(727) 478-2651** or email us on sales@ardenphotonics.com

For Rest of World sales enquiries call **+44 (0)121 733 7721** or email us on sales@ardenphotonics.com

Iss 04 Aug 16

Manufactured by
Arden Photonics Ltd

Arden Photonics Ltd,
Royston House, 267 Cranmore Boulevard,
Shirley, Solihull, B90 4QT, UK
+44 (0)121 733 7721

Arden Photonics, LLC,
4500 140th Avenue North, Suite 101,
Clearwater, FL 33762, USA
+1 (727)478-2651

www.ardenphotonics.com
enquiries@ardenphotonics.com